



Contribution ID: 92

Type: Oral

Measuring and Simulating Capture γ -Ray Spectra using the RPI γ -Multiplicity Detector

Thursday, 20 July 2023 17:30 (15 minutes)

Accurate modeling of neutron induced capture γ -production is essential for reactor and shielding calculations, understanding γ -heating in critical systems, and non-proliferation applications. To determine the accuracy of nuclear data evaluations and simulation tools used to transport capture γ -cascades, the 16-segment γ -multiplicity NaI(Tl) detector at the Rensselaer Polytechnic Institute (RPI) Gaertner Linear Accelerator Center (LINAC) has been upgraded to measure capture γ -ray spectra and multiplicity as a function of energy. Several samples including Fe, Mn, Co, Ta, and $^{235,238}\text{U}$ have been measured using the time-of-flight (TOF) method for incident neutrons in the low-energy region from 0.01 – 100 eV. A new method has been developed to model the event-by-event capture γ -cascade energy deposition in the detector array using DICEBOX and a modified version of MCNP-6.2. The method has been validated using ^{22}Na and ^{60}Co coincidence sources and the well-studied thermal $^{56}\text{Fe}(n,\gamma)$ capture γ -ray intensities. Additional measured samples will be used for further validation and analysis. The new modeling capabilities coupled with measured γ -ray spectra can be used to test transport codes and nuclear data evaluations of capture γ -rays used to simulate experimental results.

Primary authors: COOK, Katelyn (Rensselaer Polytechnic Institute); Dr DANON, Yaron (RPI); Dr LEWIS, Amanda (Naval Nuclear Laboratory)

Co-authors: Dr RAPP, Michael (Naval Nuclear Laboratory); Dr DASKALAKIS, Adam (Naval Nuclear Laboratory); Dr BARRY, Devin (Naval Nuclear Laboratory); BRAIN, Peter (RPI); Dr FRITZ, Dominik (RPI); GOLAS, Alec (RPI); Dr NEY, Adam (RPI); SINGH, Sukhjinder (RPI); WANG, Benjamin

Presenter: COOK, Katelyn (Rensselaer Polytechnic Institute)

Session Classification: Session 13A

Track Classification: Nuclear Reactions